



Government of West Bengal
OFFICE OF THE PRINCIPAL
MAHARAJA JITENDRA NARAYAN MEDICAL COLLEGE & HOSPITAL
(Previously Cooch Behar Govt Medical College & Hospital)
Vivekananda Street, Pilkhana, Cooch Behar-736101



Tel: 75010 37888

Email: principalmjnmch@gmail.com

Web: www.mjnmch.ac.in

Memo No.: MJNMC/Prin/662/2022

Date 11/04/2022

TENDER NOTICE

The Principal, Maharaja Jitendra Narayan Medical College & Hospital Cooch Behar hereby invites Tender from the bonafide Firms/Supplier for supply & installation (Including Fitting, Fixing with installation Charge) of different items including equipments, to be installed in the Department of Microbiology (2nd Call), MJN MCH premises, the specification are mentioned in attached sheet. In view of above, the undersigned invites bid in separate sealed envelope in official pad (technical bid and financial bid separately) from interested parties with the terms and conditions noted below:

The rate should be quoted both in figure and words of each item excluding of GST. GST % should be mentioned separately at prescribed column.

Earnest money : Rupees Seven Thousand (Rs. 7,000.00/-) only as Demand Draft/NEFT from any nationalized Bank in favour of The Principal, MJN MCH (A/C No.-919010029879092, IFSC- UTIB0000396, AXIS BANK- Cooch Behar Branch), payable at Cooch Behar.

The Security Deposit will be forfeited under following conditions:

- Bad performance & Bad Quality of the Items.

The last date of submission of tender is **20/04/2022** and the technical bid will be opened on the next day, on 21/04/2022 at 12.00 noon in the office of the undersigned. Financial Bid of only successful bidders will be opened on the same day. The said tender paper documents regarding the terms and conditions will be available in office of the undersigned.

The bid documents are not transferable.

The Tenders are to be accompanied with the following papers / documents:

- Bank Draft/NEFT Receipt indicating deposit towards payment of EMD.
- Copy of Valid Trade License (Nature of Business)
- Copy of PAN Card.
- Copy of updated Professional Tax Certificate
- Copy of GST Registration Certificate (Regular)
- Copy of updated Income Tax Certificate.

Principal
M.J.N. Medical College & Hospital
Cooch Behar

Date 11/04/2022

Memo No.: MJNMC/Prin/662/2022

Copy forwarded for information and necessary action please for displaying on the notice board M.J.N. Medical College & Hospital Cooch Behar - 736101 (W.B)

- 1) The Sabhadhipati, Cooch Behar Zilla Parishad, Cooch Behar
- 2) The District Magistrate, Cooch Behar
- 3) The MSVP, MJN MCH, Cooch Behar
- 4) The Board of Administrator/Chairman, Cooch Behar Municipality, Cooch Behar
- 5) The D.I.O., NIC, Cooch Behar, with request to upload the notice in www.coochbehar.gov.in
- 6) The Accounts Officer, MJN MCH, Cooch Behar
- 7) The IT Cell, MJN MCH with request to upload the notice in www.mjnmch.ac.in
- 8) Notice Board of this Office
- 9) Office copy

Principal
M.J.N. Medical College & Hospital
Cooch Behar

Principal
M.J.N. Medical College & Hospital
Cooch Behar - 736101 (W.B)

LIST OF ITEMS

SI No	Name of the Item	Indicative Specification	Quantity
1	Multichannel Micropipette	120-1200ul	2
2	Micropipette Single Channel	100-1000ul	1
3	Micropipette Single Channel	10-100ul	1
4	BOD Incubator	Volume: 68 Litres ; Capacity: 2.4 cu. Ft. Ambient condition: 0-50° C Operating/produccible/achievable temperature range in BOD incubator: +2° C to +60° C Temperature uniformity +/- 0.5° C at 20° C BOD Outer Layer: Rust, corrosion proof Lock and Buzzer alarm for door: Yes Calibration and operating system: Yes	1
5	Centrifuge	Speed Regulation: Step-less speed regulator Locking System: Safety Lid interlock to prevent cover opening during centrifugation) Rotors : Swing out / Angle rotors and reduction adaptors Max Speed: 4400 rpm Max RCF: 2350 g Max Capacity: 200ml approx. Countdown Timer: Digital Speed Indicator: Digital	2
6	Slit air sampler System	Flow rates: 30-60-90-100-120 l/mi (ISO/TC 209) with anemometer 1-2000 litres (and more with manual control) 335X135X175 MM (L X W X H)	1